

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 46D2146160	(X3) Date Survey Completed 01/26/2023
Name of Provider or Supplier Vivida Dermatology	Street Address, City, State 1490 E Foremaster Dr #150, Saint George, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview with the laboratory manager, the laboratory failed to establish a written policy to assess personnel competency for mid-level practitioners who conducted (potassium hydroxide) KOH testing since the last survey on February 4, 2020. Findings include: 1. Review of laboratory policies and procedures revealed the laboratory failed to establish and perform competency assessment for two of two testing personnel who were mid-level practitioners conducting KOH testing. 2. On January 26, 2023 at approximately 10:00 AM, the laboratory director confirmed there was no competency assessment policy established or performed for mid-level practitioners conducting KOH testing.</p>