

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 46D2170970	(X3) Date Survey Completed 09/14/2021
Name of Provider or Supplier Sequel Youth Services Of Lava Heights Academy	Street Address, City, State 730 Spring Drive, Toquerville, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D1002	<p>REPORTING OF SARS-CoV-2 TEST RESULTS</p> <p>During the Public Health Emergency, as defined in 400.200 of this chapter, each laboratory that performs a test that is intended to detect SARS-CoV-2 or to diagnose a possible case of COVID-19 (hereinafter referred to as a "SARS-CoV-2 test") must report SARS-CoV-2 test results to the Secretary in such form and manner, and at such timing and frequency, as the Secretary may prescribe.</p> <p>This CONDITION is not met as evidenced by: Based on record review and interview the laboratory failed to report SARS-Co-V-2 negative test results from January 2021 through August 2021. Findings include: 1. Email response to survey request received 9/14/2021 from laboratory director indicated only positive SARS-CoV-2 test results were reported to Southwest Utah Public Health Department. 2. Laboratory procedure failed to include instructions for reporting negative SARS-CoV-2 test results to the state health department. 3. The laboratory performed approximately 30 SARS-CoV-2 tests using the Abbott BinaxNOW COVID-19 Antigen test annually. 4. The laboratory director confirmed the findings by phone call on 09/14/2021 at 10:30 AM.</p>