

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 46D2171278	(X3) Date Survey Completed 06/12/2025
Name of Provider or Supplier Smart Skin Dermatology	Street Address, City, State 3200 Clubhouse Drive Ste 100, Lehi, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on direct observation and interview with the Office Manager, the laboratory failed to ensure that the staining reagents were not used past the expiration date. The laboratory performs approximately 144 histopathology tests annually. Findings Include: 1. Direct observation of reagents revealed the Acetone, lot number 152630, had an expiration date of 2024-05-31, Eosin-Y Vintage, lot number 183089, had an expiration date of 2025-04-30, Gill 3 Hematoxylin, lot number 151413, had an expiration date of 2023-12-31, Eosin-Y Alcoholic 0.25%, lot number 149182, had an expiration date of 2024-05-31, and 100% Reagent Alcohol, lot number 152460, had an expiration date of 2024-06-30 at 12:55 PM on 06/12/2025. 2. Interview with the Office Manger at 12:55 PM on 06/12/2025 confirmed the staining reagents were used for patient testing past the expiration date.</p>