

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 46D2193220	(X3) Date Survey Completed 01/23/2024
Name of Provider or Supplier Tooele Dermatology	Street Address, City, State 2321 N 400 E, #200, Tooele, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on direct observation and interview with histotechnician, the laboratory failed to ensure that the Potassium Hydroxide reagent was not used past the expiration date. The laboratory performs approximately 15-20 mycology tests annually. Findings Include: 1. Direct observation of Potassium Hydroxide on 1/23/2024 at 12:00 PM, revealed the Potassium Hydroxide, lot number 0043, had an expiration date of 2022-2-12. 2. Interview with the histotechnician at 12:00 PM on 1/23/2024, confirmed the Potassium Hydroxide was used for patient testing past the expiration date.</p>