

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 46D2194863	(X3) Date Survey Completed 04/25/2025
Name of Provider or Supplier Central Utah Dermatology	Street Address, City, State 861 N Main, Richfield, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on direct observation and interview with the lab manager, the laboratory failed to ensure that tissue marking dyes were not used past their expiration date. The laboratory performed approximately 1,200 histopathology tests annually. Findings Include: 1. Direct observation of tissue marking dyes on 04/25/2025 at 2:45 PM revealed that 2 of 7 tissue marking dyes had expired. The blue CDI tissue marking dye with lot number 21322 had an expiration date of 11/30/2023. The green CDI tissue marking dye with lot number 220110 had an expiration date of 01/31/2024. 2. Interview with the lab manager on 04/25/2025 at 2:45 PM confirmed the tissue marking dyes were being utilized for patient testing past their expiration date.</p>