

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  46D2297883	<b>(X3) Date Survey Completed</b>  04/28/2025
<b>Name of Provider or Supplier</b>  Vasectomy Verify Llc	<b>Street Address, City, State</b>  130 Snyders Lane, Coalville, UT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D6046</b>	<p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(8)</p> <p>(b)(8) Evaluating the competency of all testing personnel and assuring that the staff maintain their competency to perform test procedures and report test results promptly, accurately and proficiently. The procedures for evaluation of the competency of the staff must include, but are not limited to--</p> <p>This STANDARD is not met as evidenced by:</p> <ul style="list-style-type: none"> <li>_ Based on lack of documentation, and interview with Testing Personnel 1 (TP1), the laboratory failed to evaluate the competency of 2 of 2 testing personnel performing qualitative semen analysis. The laboratory performs approximately 3,500 tests a year.</li> <li>_ Findings include: 1. Record review revealed the laboratory failed to document and perform competency assessment for 2 of 2 testing personnel performing qualitative semen analysis. 2. In an interview on 4/38/2025 at approximately 2:06 PM, TP1 confirmed that the laboratory failed to document and perform competency assessment for 2 of 2 testing personnel.</li> </ul>