Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 47D2198043	(X3) Date Survey Completed 09/28/2021
Name of Provider or Supplier Brookdale Fillmore Pond	**Street Address, City, State** 300 Village Ln, Bennington, VT	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D1002	REPORTING OF SARS-CoV-2 TEST RESULTS During the Public Health Emergency, as defined in 400.200 of this chapter, each laboratory that performs a test that is intended to detect SARS-CoV-2 or to diagnose a possible case of COVID-19 (hereinafter referred to as a "SARS-CoV-2 test") must report SARS-CoV-2 test results to the Secretary in such form and manner, and at such timing and frequency, as the Secretary may prescribe. This CONDITION is not met as evidenced by: Based on staff interview, the laboratory (lab) failed to provide documentation demonstrating COVID-19 test results performed from October 2020 to February 2021. Findings include: Interview on 9/28/2021 via electronic mail with the laboratory director (LD) revealed the lab had performed an unspecified number of tests for COVID-19 but due to key personnel changes the test records could not be located. The LD further revealed the COVID-19 test results were not reported to Vermont Department of Health (VDH) and testing ceased in February 2021. Review of CLIA records on 9/28/2021 revealed testing for COVID-19 began in October 2020.