

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 48D2118495	(X3) Date Survey Completed 05/23/2019
Name of Provider or Supplier St Croix Cancer Specialists	Street Address, City, State 3018 Orange Grove, Christiansted, VI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on lack of documentation and confirmation by the laboratory director, the laboratory failed to follow its polices and procedures for assessing the competency of three of four testing personnel responsible for patient testing using the Coulter AcT 5 hematology analyzer.</p>