

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 49D0016288	(X3) Date Survey Completed 10/08/2019
Name of Provider or Supplier Vpi And Student Health Services Laboratory	Street Address, City, State Mccomas Hall, Virginia Tech, 895 Washington St, Sw, Blacksburg, VA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA Recertification survey was conducted at the VPI and Student Health Services Laboratory on October 8, 2019 by the Virginia Department of Health's Office of Licensure and Certification. The laboratory was surveyed under 42 CFR part 493 CLIA Requirements. Specific deficiencies cited are as follows:
D5807	<p>TEST REPORT CFR(s): 493.1291(d)</p> <p>Pertinent "reference intervals" or "normal" values, as determined by the laboratory performing the tests, must be available to the authorized person who ordered the tests and, if applicable, the individual responsible for using the test results.</p> <p>This STANDARD is not met as evidenced by: Based on a review of patient test reports, lack of documentation, and interview with the technical consultant, the laboratory's Mediat Electronic Health Record (EHR) final patient report failed to contain reference intervals or normal values for two (2) urine colony counts, one (1) urine microscopy, and one (1) vaginal wet preparation examination at the date of survey on October 8, 2019. Findings include: 1. Review of patient test reports generated from the Mediat EHR revealed lack of documentation of reference intervals or normal values for 2 urine colony counts, 1 urine microscopy and 1 vaginal wet preparation examination. 2. An interview with the technical consultant at approximately 2:45 PM confirmed the findings.</p>