

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  49D0667592	<b>(X3) Date Survey Completed</b>  09/20/2018
<b>Name of Provider or Supplier</b>  Centra Medical Group - Amherst	<b>Street Address, City, State</b>  115 Ambriar Court, Amherst, VA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	An announced CLIA Recertification survey was conducted at Centra Medical Group-Amherst on September 20, 2018 by the Virginia Department of Health's Office of Licensure and Certification. The laboratory was surveyed under 42 CFR part 493 CLIA Requirements. Specific deficiencies cited are as follows:
<b>D6053</b>	<p><b>TECHNICAL CONSULTANT RESPONSIBILITIES</b> CFR(s): 493.1413(b)(9)</p> <p>The technical consultant is responsible for evaluating and documenting the performance of individuals responsible for moderate complexity testing at least semiannually during the first year the individual tests patient specimens.</p> <p>This STANDARD is not met as evidenced by: Based on review of the Centers for Medicare and Medicaid Services Laboratory Personnel Report form (CMS 209), testing personnel (TP) A competency records, and interviews, the technical consultant failed to perform the semi-annual competency evaluation for one (1) of one (1) new testing personnel in 2018. (see Attached Personnel Code Sheet). Findings include: 1. Review of the CMS 209: Laboratory Personnel form revealed that the laboratory director also performs the duties of technical consultant. 2. Review of TP A records revealed that TP A was hired in October 2017 and that there was no documentation of a semi-annual competency performed at the date of survey. The inspector requested to view aforementioned documentation. The documentation was not available for review. 3. An interview with the technical consultant at approximately 12:00 PM confirmed the above-listed findings.</p>