

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 49D0688909	(X3) Date Survey Completed 07/05/2018
Name of Provider or Supplier Dermatology Consultants Inc	Street Address, City, State 1330 Oak Lane - Suite 101, Lynchburg, VA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA Recertification survey was conducted at the Dermatology Consultants on July 5, 2018 by the Virginia Department of Health's Office of Licensure and Certification. The laboratory was surveyed under 42 CFR part 493 CLIA Requirements. Specific deficiencies cited are as follows:
D5401	<p>PROCEDURE MANUAL CFR(s): 493.1251(a)</p> <p>A written procedures manual for all tests, assays, and examinations performed by the laboratory must be available to, and followed by, laboratory personnel. Textbooks may supplement but not replace the laboratory's written procedures for testing or examining specimens.</p> <p>This STANDARD is not met as evidenced by: Based on the tour of the laboratory, review of policy and interviews, the laboratory did not follow the written policy for monitoring the Olympus microscope (Serial number 211956) at the date of survey on July 5, 2018. Findings include: 1. Tour of the laboratory revealed that the laboratory performs Potassium Hydroxide (KOH), scabies and Tzank smear microscopic examinations using the Olympus microscope (Serial number 211956). The inspector noted the visible layer of dust on the oculars and stage of the microscope. 2. Review of the policy manual revealed the following statement: "Quality control policies and documentation: The following equipment must be monitored: Microscope: Cleaning of stage and ocular eye piece." The inspector requested documentation of the cleaning of the microscope. There was no documentation available for review. 3. An interview with the laboratory director at approximately 12:00 PM confirmed that the microscope had not been monitored for cleaning according to the written policy at the date of survey.</p>