

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 49D0890671	(X3) Date Survey Completed 07/08/2026
Name of Provider or Supplier Shenandoah Oncology, Pc	Street Address, City, State 400 Campus Boulevard - Suite 100, Winchester, VA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at Shenandoah Oncology on July 7-8, 2026 by the Virginia Department of Health's Office of Licensure and Certification. The laboratory was surveyed under 42 CFR part 493 CLIA Regulations. The specific deficiency cited is as follows:
D6046	TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(8) (b)(8) Evaluating the competency of all testing personnel and assuring that the staff maintain their competency to perform test procedures and report test results promptly, accurately and proficiently. The procedures for evaluation of the competency of the staff must include, but are not limited to-- This STANDARD is not met as evidenced by: Based on a review of the Centers for Medicare and Medicaid Services Laboratory Personnel Report form (CMS 209), personnel records, lack of documentation and an interview, the technical consultant (TC) failed to assess the annual Vitros XT3400 Chemistry analyzer competency for one (1) of five (5) testing personnel (TP). The findings include: 1. Review of the CMS 209 revealed 5 TP identified as responsible for moderate complexity Chemistry patient testing. 2. Review of the laboratory's personnel records revealed a lack of the annual Vitros XT3400 competency assessment for Testing Personnel C in 2025 (initial competency 06/10/2024, semi-annual competency 12/12/2024). See testing personnel code sheet. The surveyor requested to review TP C's 2025 Vitros XT3400 competency. The laboratory provided no documentation for review. 3. In an exit interview with the laboratory director, technical consultant and clinical manager at 11:30 AM on July 8, 2026, the above findings were confirmed.