

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 49D0978157	(X3) Date Survey Completed 11/09/2020
Name of Provider or Supplier A Capital Womens Health Clinic	Street Address, City, State 1511 Starling Drive, Richmond, VA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA Recertification on-site survey was conducted at A Capital Women's Health Center on November 9, 2020 by the Virginia Department of Health's Office of Licensure and Certification. The laboratory was surveyed under 42 CFR part 493 CLIA Requirements. The initial contact with laboratory was on October 14, 2020 with off-site record review of documentation and a follow-up phone conference on November 4, 2020. Specific deficiencies cited are as follows:
D2159	<p>ABO GROUP AND D(RHO) TYPING CFR(s): 493.859(d)</p> <p>Failure to return proficiency testing results to the proficiency testing program within the time frame specified by the program is unsatisfactory performance and results in a score of 0 for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on the review of proficiency testing (PT) records, CASPER Report 0096D and interview, the laboratory failed to submit the PT results within the specified time frame for the ABO/ RHO (Blood group typing) for one (1) of 3 events in 2019. Findings include: 1. Review of the American Proficiency Institute (API) PT records for all testing events in 2019 revealed the following: RHO (D) Type Event 1- 100% RHO (D) Type Event 2- 100% RHO (D) Type Event 3- 0% *Failure to Participate 2. Review of the CASPER Report 0096D revealed the following event with results less than 100%: 2019 Event 3- ABO/RHO = 0% and D (RHO) = 0% 3. An interview with the site administrator on November 9, 2020 at approximately 2:00 PM confirmed the findings.</p>