

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 49D0990271	(X3) Date Survey Completed 09/25/2019
Name of Provider or Supplier New River Dermatology	Street Address, City, State 2617 Sheffield Drive, Blacksburg, VA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA Recertification survey was conducted at the New River Dermatology on September 25, 2019 by the Virginia Department of Health's Office of Licensure and Certification. The laboratory was surveyed under 42 CFR part 493 CLIA Requirements. Specific deficiencies cited are as follows:
D6120	<p>TECHNICAL SUPERVISOR RESPONSIBILITIES CFR(s): 493.1451(b)(7)(8)</p> <p>(7) The technical supervisor is responsible for identifying training needs and assuring that each individual performing tests receives regular in-service training and education appropriate for the type and complexity of the laboratory services performed; (8) Evaluating the competency of all testing personnel and assuring that the staff maintain their competency to perform test procedures and report test results promptly, accurately and proficiently.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the Laboratory Personnel Report form (CMS 209), testing personnel (TP) records, lack of documentation, and an interview with the technical supervisor (TS), the TS failed to assess annual competency for one (1) of 1 TP in 2018. Findings include: 1. Review of the CMS Form 209: Laboratory Personnel Report revealed that the laboratory director serves as the TS and that there is 1 TP performing grossing of patient tissue samples in the laboratory. 2. Review of TP A records revealed a lack of documentation of an annual competency assessment in 2018 (See Testing Personnel Code Sheet.) 3. An interview with the TS at approximately 12:30 PM confirmed the findings.</p>