

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  49D1048430	<b>(X3) Date Survey Completed</b>  09/23/2025
<b>Name of Provider or Supplier</b>  Christine S Rausch Md, Pc	<b>Street Address, City, State</b>  2510 Gaskins Road, Richmond, VA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	An announced CLIA recertification survey was conducted at Christine S. Rausch, MD, PC (Elevate Dermatology and Plastic Surgery) on September 23, 2025 by the Virginia Department of Health's Office of Licensure and Certification. The laboratory was surveyed under 42 CFR part 493 CLIA Regulations. Specific deficiency cited is as follows:
<b>D5805</b>	<p>TEST REPORT CFR(s): 493.1291(c)</p> <p>(c) The test report must indicate the following: (c)(1) For positive patient identification, either the patient's name and identification number, or a unique patient identifier and identification number. (c)(2) The name and address of the laboratory location where the test was performed. (c)(3) The test report date. (c)(4) The test performed. (c)(5) Specimen source, when appropriate. (c)(6) The test result and, if applicable, the units of measurement or interpretation, or both. (c)(7) Any information regarding the condition and disposition of specimens that do not meet the laboratory's criteria for acceptability.</p> <p>This STANDARD is not met as evidenced by: Based on a review of seven randomly selected patient histology slides and Mohs surgical maps, lack of documentation and an interview, the laboratory failed to ensure the laboratory's address was included on seven of seven selected Mohs surgical maps on the date of the survey on September 23, 2025. The findings include: 1. The surveyor randomly selected seven patients for documentation comparison of the histology slides to the Mohs final reports. The laboratory provided the surveyor the patient slides and corresponding Mohs surgical maps for the seven patients. 2. During a review of the slides and maps, the surveyor observed seven of seven Mohs surgical maps lacked documentation of the laboratory's address (case numbers M23-1466, M24-545, M24-1123, M24-1606, M25-023, M25-690, and M25-1226). During an interview with the Mohs technician on September 23, 2025 at 11:45 AM, the surveyor</p>

inquired what the laboratory considered the final report. The Mohs technician stated the Mohs maps were the final report. 3. In an exit interview with the Mohs technician on September 23, 2025 at 1:00 PM, the findings were confirmed.