

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 49D1091651	(X3) Date Survey Completed 09/26/2024
Name of Provider or Supplier Evms Dermatology- Andrews Hall	Street Address, City, State 721 Fairfax Avenue - Suite 200, Norfolk, VA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA recertification survey was conducted at EVMS Dermatology- Andrews Hall on September 26, 2024 by the Virginia Department of Health's Office of Licensure and Certification. The laboratory was surveyed under 42 CFR part 493 CLIA Requirements. Specific deficiency cited is as follows:
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the Centers for Medicare and Medicaid Services Laboratory Personnel Report Form (CMS 209), mycology proficiency testing (PT) records, and an interview, the laboratory failed to verify testing personnel (TP) accuracy of Potassium Hydroxide (KOH) mycology microscopy for two (2) of 2 PT events in calendar year 2024 as noted on the date of the inspection, September 26, 2024. Findings include: 1. Review of the laboratory's CMS 209 personnel form revealed that four TP qualified to perform patient KOH microscopy examinations. (See Testing Personnel Code Sheet.) 2. Review of the laboratory's year to date 2024 American Proficiency Institute (API) Microscopy PT documentation, a total of 2 events, revealed the laboratory utilized PT to verify TP KOH microscopy accuracy. A review of the API PT reports revealed the laboratory failed accuracy verification year to date by receiving the following unsatisfactory scores: 2024 1st Event: 50% Unsatisfactory Scoring; failed for challenge KOH--02; 2024 2nd Event: 50% Unsatisfactory Scoring; failed for challenge KOH--03. The API report noted the laboratory as unsuccessful long term. 3. An interview with the practice manager, lab director, and TP on 9/26/24 at 11:30 AM confirmed the above findings.</p>