

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 49D2025069	(X3) Date Survey Completed 09/29/2021
Name of Provider or Supplier Absolute Dermatology Pc	Street Address, City, State 5207 Hickory Park Drive - Suite A, Glen Allen, VA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced CLIA Recertification survey was conducted at Absolute Dermatology on September 29, 2021 by the Virginia Department of Health's Office of Licensure and Certification. The laboratory was surveyed under 42 CFR part 493 CLIA Requirements. Specific deficiencies cited are as follows:
D6128	<p>TECHNICAL SUPERVISOR RESPONSIBILITIES CFR(s): 493.1451(b)(9)</p> <p>The technical supervisor is responsible for evaluating and documenting the performance of individuals responsible for high complexity testing at least annually after the first year, unless test methodology or instrumentation changes, in which case, prior to reporting patient test results, the individual's performance must be reevaluated to include the use of the new test methodology or instrumentation.</p> <p>This STANDARD is not met as evidenced by: Based on a review of the Laboratory Personnel Report form (CMS 209), testing personnel (TP) files, lack of documentation, and an interview, the technical supervisor (TS) failed to assess annual competency for one (1) of two (2) testing personnel in 2020. Findings include: 1. Review of the CMS Form 209 Laboratory Personnel Report revealed that the laboratory director serves as the TS, and TP A performed high complexity testing in 2020. (See Personnel Code Sheet.) 2. Review of TP files revealed lack of documentation of an annual competency assessment for TP A in the calendar year 2020. 3. An exit interview with TP A and the practice manager at approximately 2 PM on 09/29/21 confirmed the findings.</p>