Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number** 49D2043119	**(X3) Date Survey Completed** 06/17/2024
Name of Provider or Supplier Laboratory Corporation Of America Holdings	**Street Address, City, State** 21 Highland Avenue, Se - Suite 200, Roanoke, VA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	**Summary Statement of Deficiencies**
D0000	An announced CLIA Recertification survey was conducted at the Laboratory Corporation of America Holdings on 06/17/24 by the Virginia Department of Health's Office of Licensure and Certification. The laboratory was surveyed under 42 CFR part 493 CLIA Requirements. Specific deficiencies cited are as follows:
D2007	TESTING OF PROFICIENCY TESTING SAMPLES CFR(s): 493.801(b)(1) The samples must be examined or tested with the laboratory's regular patient workload by personnel who routinely perform the testing in the laboratory, using the laboratory's routine methods This STANDARD is not met as evidenced by: Based on the review of the Laboratory Personnel Report Form (CLIA) (CMS-209 Form), proficiency testing (PT) records, and an interview, the lab failed to rotate five of five chemistry and five of five hematology PT events among two testing personnel (TP) for the ten of ten events reviewed between September 29, 2022 up to the date of survey on 06/17/24. Findings include: 1. Review of the CMS-209 personnel form revealed two TP performing serum human chorionic gonadotropin (HcG) chemistry and hematology microscopy test procedures in the calendar year 2022, 2023 and up to the date of survey on 06/17/24. 2. Review of the American Proficiency Institute (API) PT records revealed that the same TP performed the following 10 events (TP-A): 2022- Chemistry, Serum HcG event 3, 2022- Hematology Microscopy event 3. 2023- Chemistry, Serum HcG event 1, 2 and 3, 2023- Hematology Microscopy event 1, 2 and 3. 2024- Chemistry, Serum HcG event 1, 2024- Hematology Microscopy event 1. In an interview with the technical consultant on 06/17/24 at 10 AM, they stated "Testing personnel B works at a affiliated site and performs PT there for the same modules. Moving forward, TP B will be rotated with this CLIA number." 3. An exit interview with the technical consultant on 06/17/24 at 11 AM confirmed the findings.