

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 49D2080254	(X3) Date Survey Completed 10/28/2020
Name of Provider or Supplier Tamjidi Skin Institute	Street Address, City, State 8100 Boone Boulevard - Suite 300, Vienna, VA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced on-site CLIA recertification survey was conducted at Tamjidi Skin Institute on October 28, 2020 by the Virginia Department of Health's Office of Licensure and Certification. The survey included an entrance interview on 09/25/2020 and virtual record review conducted on 10/20/2020. The laboratory was surveyed under 42 CFR part 493 CLIA Regulations. The specific deficiency is as follows:
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on a laboratory tour, review of procedures, Quality Assurance documents and interviews, the laboratory failed to ensure the Avantik Eosin Working Solution, stored in the laboratory for use, was within the manufacturer's stated expiration dates. Findings include: 1. During a laboratory tour, at approximately 9:00 AM, the surveyor noted a bottle of Avantik Eosin Working Solution (lot number H057-21) in the flammable cabinet with a hand written receive date of 12/27/18 and manufacturer stamped expiration date of 3/6/20. The inspector noted a written "opened" date of 4/21/2020 with approximately one-third of the solution remaining. No additional bottles of Eosin were observed in the storage cabinet or logged in the reagent log as in use. 2. Review of the laboratory's procedure manual revealed a "Monthly Quality Assurance Checklist" that stated, "All reagents, controls, kits, etc that exceeded their expiration date were discarded." The checklist was completed as "(Y)es" from April 2020 to September 2020. 3. Review of the laboratory's reagent log revealed the Avantik Eosin Working Solution lot H057-21 expiration date 3/6/20 was documented as opened/in use on 4/21/2020. 4. Review of the laboratory's Moth's log revealed one-hundred thirty-four (134) patient Mohs slides were prepared and stained with Hematoxylin and</p>

Eosin stain from 4/22/2020 until 10/26/2020. 5. In an interview with the Clinical Manager at approximately 10:15 AM on 10/28/2020, the above findings were confirmed.