

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 49D2169537	(X3) Date Survey Completed 06/25/2020
Name of Provider or Supplier Northern Virginia Carenow Urgent Care Llc	Street Address, City, State 1380 N Parham Rd, Richmond, VA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An unannounced off-site CLIA complaint investigation was conducted at Bettermed Urgent Care-Regency from May 27, 2020 up to June 25, 2020 by the Virginia Department of Health's Office of Licensure and Certification. The laboratory was surveyed under 42 CFR Part 493 CLIA requirements. Specific deficiencies cited are as follows:
D3009	<p>FACILITIES CFR(s): 493.1101(c)</p> <p>The laboratory must be in compliance with applicable Federal, State, and local laboratory requirements.</p> <p>This STANDARD is not met as evidenced by: Based on review of the Virginia State Board of Health Regulations and the Federal Coronavirus Aid, Relief, and Economic Security (CARES) Act, telephone interviews with the Director of Clinical Operations and the Chief Operating Officer (COO), and review of available patient test logs, the laboratory failed to report positive patient test results for the 2019 novel Coronavirus (COVID-19) antibody in accordance with State and Federal requirements for thirty-two (32) of six hundred and sixty (660) patients from May 5, 2020 through May 28, 2020. Findings include: 1. Review of the Commonwealth of Virginia State Board of Health November 2018, Regulations for Disease Reporting and Control (page 10) revealed the following statements: "12 VAC 5-90-80. Lists of diseases that shall be reported. A. Reportable disease list. The board declares suspected or confirmed cases of the following named diseases, toxic effects, and conditions to be reportable by the persons enumerated in 12VAC5-90-90. Conditions identified by an asterisk (*) require immediate communication to the local health department by the most rapid means available upon suspicion or confirmation, as defined in subsection C of this section. *Coronavirus infection, severe C. Reportable diseases requiring rapid communication. Certain of the diseases in the list of reportable diseases because of their extremely contagious nature, potential for</p>

greater harm, or availability of a specific intervention that must be administered in a timely manner require immediate identification and control. Reporting of persons confirmed or suspected of having these diseases, listed in this subsection, shall be made immediately by the most rapid means available, preferably by telephone to the local health department (page 13). Coronavirus infection, severe" 2. Review of the Coronavirus Aid, Relief, and Economic Security (CARES) Act Public Law 116-136, 18115(a) revealed the following statements: "requires every laboratory that performs or analyzes a test that is intended to detect SARS-CoV-2 or to diagnose a possible case of COVID-19 to report the results from each such test to the Secretary of the Department of Health and Human Services (HHS). This document outlines the requirements for data submission to HHS as authorized under this law. In an effort to receive these data in the most efficient and effective manner, the Secretary is requiring that all data be reported through existing public health data reporting methods, described below. As a guiding principle, data should be sent to state or local public health departments using existing reporting channels (in accordance with state law or policies) to ensure rapid initiation of case investigations by those departments, concurrent to laboratory results being shared with an ordering provider, or patient as applicable." 3. An interview with the Director of Clinical Operations and the COO on May 28, 2020 at approximately 1 PM revealed the laboratory utilized the non FDA approved Corona Check (Innovita) COVID-19 IgG/IgM Antibody test method to detect IgG/IgM antibodies in patients to the 2019 novel Coronavirus (COVID-19) from 5/06/20 through 5/28/20. In addition, lack of communication to local and State health departments of patient positive test results for COVID-19 IgG/IgM antibody for the specified timeframe in accordance with the aforementioned State and Federal regulations. 4. Review of available patient test logs revealed the following patients reported as positive for COVID-19 IgG/IgM Antibody: 5/5/20- 1 patient, 5/7/20- 2 patients, 5/8/20- 1 patient, 5/9/20- 2 patients, 5/10/20- 2 patients, 5/11/20- 1 patient, 5/12/20- 2 patients, 5/13/20- 4 patients, 5/14/20- 5 patients, 5/16/20- 1 patient, 5/18/20- 2 patients, 5/20/20- 2 patients, 5/21/20- 1 patient, 5/22/20- 3 patients, 5/24/20- 1 patient, 5/26/20- 1 patient and, 5/27/20- 1 patient. Total of 32 patients. 5. An interview with the Director of Clinical Operations and the COO on June 25, 2020 at approximately 11 AM confirmed the findings.

D5400

ANALYTIC SYSTEMS
CFR(s): 493.1250

Each laboratory that performs nonwaived testing must meet the applicable analytic systems requirements in 493.1251 through 493.1283, unless HHS approves a procedure, specified in Appendix C of the State Operations Manual (CMS Pub.7), that provides equivalent quality testing. The laboratory must monitor and evaluate the overall quality of the analytic systems and correct identified problems as specified in 493.1289 for each specialty and subspecialty of testing performed.

This CONDITION is not met as evidenced by:
Based on interviews, Food and Drug Administration's (FDA) Emergency Use Authorizations (EUA), review of manufacturer's package insert, policies/procedures, available logs of patient testing, quality control (QC), temperature monitoring, and lack of documentation, the laboratory failed to: 1. document monitoring temperatures to ensure proper storage of non FDA approved Corona Check 2019-nCoV Ab Test Colloidal Gold rapid test kit reagents from 5/5/20 to 5/31/20; 2. evaluate and verify the performance specifications of a non FDA approved COVID-19 IgG/IgM test method prior to reporting six hundred sixty (660) patient COVID-19 IgG/IgM

Antibody results from 5/5/20 to 5/31/20; 3. document performance of a negative and positive external controls for a non FDA approved COVID-19 IgG/IgM test method for each day of patient testing from 5/5/20 to 5/31/20. See D5413, D5423, D5449.

D5413

TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT
CFR(s): 493.1252(b)

The laboratory must define criteria for those conditions that are essential for proper storage of reagents and specimens, accurate and reliable test system operation, and test result reporting. The criteria must be consistent with the manufacturer's instructions, if provided. These conditions must be monitored and documented and, if applicable, include the following: (1) Water quality. (2) Temperature. (3) Humidity. (4) Protection of equipment and instruments from fluctuations and interruptions in electrical current that adversely affect patient test results and test reports.

This STANDARD is not met as evidenced by:

Based on interviews, review of the Food and Drug Administration's (FDA) Emergency Use Authorizations (EUA), manufacturer's package insert, policies /procedures, available test logs, and lack of documentation, the laboratory failed to document monitoring of temperatures to ensure proper storage of high complexity COVID-19 IgG/IgM antibody immunology kit reagents from May 5, 2020 to May 31, 2020. Findings include: 1. During an email interview with the the Director of Clinical Operations and Chief Operating Officer (COO) on 5/27/20 at approximately 10:10 AM, it was revealed that the laboratory had been utilizing the following COVID-19 test method (in a drive through testing setting) during the timeframe of 5/5/20 through the date of the offsite investigation: Corona Check 2019-nCoV Ab Test Colloidal Gold, manufactured in China by Innovita Tangshan Biological Technology Co. LTD. 2. Review of the FDA's published listing of COVID-19 EUA granted for SARS CoV-2 antibody testing on 5/27/20 revealed no EUA granted for the kit outlined above. The Innovita Corona Check rapid test method classification as of 5/27/20 was high complexity. 3. Review of the Innovita Corona Check package insert revealed manufacturer's instructions: "Storage and Stability: Store at 4-30 C (39.2-86 F)". 4. Review of the laboratory's approved procedure ("BetterMed POC COVID-19 Innovita Antibody Testing") revealed no instructions for monitoring temperatures for proper storage of kit/reagent. 5. Review of the available laboratory logs from 5/5/20 to 5/31 /20 revealed that 660 patient COVID-19 IgG/IgM test results were reported. No temperature monitoring logs for the Innovita kits conforming with the manufacturer's instructions were available for review. 6. During an email interview on 6/17/20, at approximately 10:30 AM, the inspectors requested clarification regarding temperature logs for the timeframe of patient testing. The COO stated: "No temperatures were recorded for the storage areas. All antibody kits were delivered to our corporate office at 3200 Rockbridge St, Suite 103 Richmond, VA 23230. Each of our test sites had a limited appointment slot of 25-35 antibody tests per day. The number of kits were delivered to get them through a 7 day period, often times it would be more than that amount so they would just receive a box of the kits instead of loose kits. Regarding distribution, the kits were distributed to each site Tuesday mornings via our courier, Orbit. At times a site manager may stop by and pick up a box or a box of kits may be distributed by someone from our corporate office that may have been going to a site that needed some. Lot numbers of the boxes or individual kits were not tracked. Boxes of kits were simply handed out". The COO stated in the follow up email interview on 6

/17/20, at approximately 12:52 PM: "BetterMed Regency has used Innovita antibody kits from 5/5/2020 - present." 7. An interview with the director of clinical operations and COO on 6/25/20, at approximately 11:00 AM, confirmed the above findings.

D5423

ESTABLISHMENT AND VERIFICATION OF PERFORMANCE
CFR(s): 493.1253(b)(2)

Each laboratory that modifies an FDA-cleared or approved test system, or introduces a test system not subject to FDA clearance or approval (including methods developed in-house and standardized methods such as text book procedures), or uses a test system in which performance specifications are not provided by the manufacturer must, before reporting patient test results, establish for each test system the performance specifications for the following performance characteristics, as applicable: (2)(i) Accuracy. (2)(ii) Precision. (2)(iii) Analytical sensitivity. (2)(iv) Analytical specificity to include interfering substances. (2)(v) Reportable range of test results for the test system. (2)(vi) Reference intervals (normal values). (2)(vii) Any other performance characteristic required for test performance.

This STANDARD is not met as evidenced by:

Based on interviews, review of the Food and Drug Administration's (FDA) Emergency Use Authorizations (EUA), available patient and quality control (QC) logs, and lack of documentation, the laboratory failed to evaluate, verify/validate the performance specifications of a high complexity COVID-19 IgG/IgM immunology test method prior to reporting six hundred sixty (660) patient COVID-19 IgG/IgM antibody results from May 5, 2020 to May 31, 2020. Findings include: 1. During an email interview with the the Director of Clinical Operations and Chief Operating Officer (COO) on 5/27/20 at approximately 10:10 AM, it was revealed that the laboratory had been utilizing the following COVID-19 test kits (in a drive through testing setting) during the timeframe of 5/5/20 through the date of the offsite investigation: Corona Check 2019-nCoV Ab Test Colloidal Gold, manufactured in China by Innovita Tangshan Biological Technology Co. LTD. 2. Review of the FDA's published listing of COVID-19 EUA's granted for SARS CoV-2 antibody testing as of 5/27/20 revealed no EUA granted for the kit outlined above. The Innovita Corona Check rapid test method classification as of 5/27/20 was high complexity. 3. During a telephone interview with the Director of Clinical Operations and COO on 5/28/20 at approximately 1 PM, the inspectors asked for a description of the laboratory's QC protocols, to review documentation of in-house validation procedures for the high complexity COVID-19 test method, and patient test logs. The COO stated at approximately 1:30 PM: "We have ordered a different manufacturer COVID rapid test that has FDA EUA and the laboratory will be moving to those kits as soon as possible. We did not perform validation studies for the test kits we have been using. We rely on the manufacturer's validation. We use the internal built in QC. We did not purchase extra quality control materials." 4. Review of patient test and QC records revealed that the laboratory reported 660 patient COVID-19 IgG/IgM antibody test results from 5/5/20 to 5/31/20 while performing zero (0) negative or positive external controls. No records of QC were available for review. No records of an approved validation study was available for review. 5. During a follow up email interview on 6/17/20 at approximately 10:30 AM, the inspectors requested clarification/documentation for negative and positive external QC and validation of accuracy. The COO stated in a follow up email response at approximately 12:52 PM: "BetterMed Regency utilized

the Innovita test kits from 5/5/20 to present. No QC materials were purchased." 6. An interview with the director of clinical operations and COO on 6/25/20, at approximately 11:00 AM, confirmed the above findings.

D5449

CONTROL PROCEDURES

CFR(s): 493.1256(d)(3)(ii)(g)

Unless CMS Approves a procedure, specified in Appendix C of the State Operations Manual (CMS Pub. 7), that provides equivalent quality testing, the laboratory must-- At least once a day patient specimens are assayed or examined perform the following for-- Each qualitative procedure, include a negative and positive control material; (g) The laboratory must document all control procedures performed.

This STANDARD is not met as evidenced by:

Based on interviews, review of the Food and Drug Administration's (FDA) Emergency Use Authorizations (EUA), available patient and quality control (QC) logs, and lack of documentation, the laboratory failed to document performance of negative and positive external quality control (QC) materials on each day of patient testing for a high complexity COVID-19 IgG/IgM test method from May 5, 2020 to May 31, 2020. Findings include: 1. During an email interview with the the Director of Clinical Operations and Chief Operating Officer (COO) on 5/27/20 at approximately 10:10 AM, it was revealed that the laboratory had been utilizing the Corona Check 2019-nCoV Ab Test Colloidal Gold rapid test method (manufactured in China by Innovita Tangshan Biological Technology Co. LTD) in a drive through laboratory setting during the timeframe of 5/5/20 through the date of the investigation. 2. Review of the FDA's published listing of COVID-19 EUA's granted for SARS CoV-2 antibody testing as of 5/27/20 revealed no EUA granted for the test method outlined above. The Innovita Corona Check rapid test method classification as of 5/27/20 was high complexity. 3. In a telephone interview with the the Director of Clinical Operations and Chief Operating Officer (COO) on 5/28/20 at approximately 1:00 PM, the inspectors asked for a description of the laboratory's QC protocols and patient test logs for the Innovita COVID-19 test method. The COO stated at approximately 1:30 PM: "We use the internal built in QC. We did not purchase extra quality control materials." 4. Review of the available patient and QC logs revealed that the facility reported 660 COVID-19 IgG/IgM results during the timeframe of 5/5/20 to 5/31/20. No record/documentation verifying negative or positive external controls on each day of patient testing was available for review. During a follow up email interview on 6/17/20 at approximately 10:30 AM, the inspectors requested clarification/documentation for negative and positive external QC on each date of patient testing. The COO stated in a follow up email response at approximately 12:52 PM: "BetterMed Regency utilized the Innovita test kits from 5/5/20 to present. No QC materials were purchased." 5. An interview with the director of clinical operations and COO on 6/25/20, at approximately 11:00 AM, confirmed the above findings.

D6076

LABORATORY DIRECTOR

CFR(s): 493.1441

The laboratory must have a director who meets the qualification requirements of 493.1443 of this subpart and provides overall management and direction in accordance with 493.1445 of this subpart.

This CONDITION is not met as evidenced by:
Based on interviews, review of the Food and Drug Administration's (FDA) Emergency Use Authorizations (EUA), available logs of patient testing/quality control (QC)/temperature monitoring, policies and procedures, Laboratory Personnel Report Form (CLIA) (CMS-209 Form), available laboratory personnel files, and lack of documentation, the laboratory director failed to: 1. ensure validation/verification performance procedures were performed for a non FDA approved COVID-19 IgG/IgM test method prior to reporting six hundred sixty (660) patient antibody results from 5/5/20 to 5/31/20; 2. ensure that QC policies were established and followed for a non FDA approved COVID-19 IgG/IgM test method from 5/5/20 to 5/31/20; 3. ensure that training and competency assessments were performed for Testing Personnel prior to testing and reporting patient results with the high complexity Corona Check COVID-19 IgG/IgM Antibody test kit. See D6086, D6093, D6102.

D6086

LABORATORY DIRECTOR RESPONSIBILITIES
CFR(s): 493.1445(e)(3)(ii)

The laboratory director must ensure that verification procedures used are adequate to determine the accuracy, precision, and other pertinent performance characteristics of the method.

This STANDARD is not met as evidenced by:
Based on interviews, review of the Food and Drug Administration's (FDA) Emergency Use Authorizations (EUA), available patient/quality control (QC) logs, and lack of documentation, the laboratory director (LD) failed to ensure validation/verification performance procedures were performed for a non FDA approved COVID-19 IgG/IgM test method prior to reporting six hundred sixty (660) patient antibody results from May 5, 2020 to May 31, 2020. Findings include: 1. During an email interview with the the Director of Clinical Operations and Chief Operating Officer (COO) on 5/27/20 at approximately 10:10 AM, it was revealed that the laboratory had been utilizing the Corona Check 2019-nCoV Ab Test Colloidal Gold rapid test kits (manufactured in China by Innovita Tangshan Biological Technology Co. LTD) in a drive through laboratory setting during the timeframe of 5/5/20 through the date of the investigation. 2. Review of the FDA's published listing of COVID-19 EUA's granted for SARS CoV-2 antibody testing as of 5/27/20 revealed no EUA granted for the test method outlined above. The Innovita Corona Check COVID-19 IgG/IgM rapid test method classification as of 5/27/20 was high complexity. 3. In a telephone interview with the the Director of Clinical Operations and Chief Operating Officer (COO) on 5/28/20 at approximately 1:00 PM, the inspectors requested a description of the laboratory's QC protocols and to review documentation of in-house validation procedures for the high complexity COVID-19 test kits utilized. The COO stated at approximately 1:30 PM: "We have ordered a different manufacturer COVID rapid test that has FDA EUA and the laboratory will be moving to those kits as soon as possible. We did not perform validation studies for the test kits we have been using. We rely on the manufacturer's validation. We use the internal built in QC. We did not purchase extra quality control materials." 4. Review of the available patient test/QC logs revealed that the laboratory reported 660 patient COVID-19 IgG/IgM test results while performing zero (0) negative or positive external controls for the timeframe of 5/5/20 to 5/31/20. No record of a LD approved validation study of accuracy was available for review. In an email interview on 6/17/20 at approximately 10:30 AM, the inspectors requested clarification regarding dates of use and records of external QC on each date of patient testing. The COO stated, in a follow up email response on 6

/17/20 at approximately 12:52 PM: "BetterMed Regency utilized the Innovita test kits from 5/5/20 to present. No QC materials were purchased." 5. An interview with the director of clinical operations and COO on 6/25/20, at approximately 11:00 AM, confirmed the above findings.

D6093

LABORATORY DIRECTOR RESPONSIBILITIES

CFR(s): 493.1445(e)(5)

The laboratory director must ensure that the quality control programs are established and maintained to assure the quality of laboratory services provided and to identify failures in quality as they occur.

This STANDARD is not met as evidenced by:

Based on interviews, review of the Food and Drug Administration's (FDA) Emergency Use Authorizations (EUA), available patient/quality control (QC) logs, policies and procedures, and lack of documentation, the laboratory director (LD) failed to ensure that QC policies were established and followed from May 5, 2020 to May 31, 2020. Findings include: 1. During an email interview with the the Director of Clinical Operations and Chief Operating Officer (COO) on 5/27/20 at approximately 10:10 AM, it was revealed that the laboratory had been utilizing the Corona Check 2019-nCoV Ab Test Colloidal Gold rapid test kits (manufactured in China by Innovita Tangshan Biological Technology Co. LTD) for a drive through laboratory test setting during the timeframe of 5/5/20 through the date of the investigation. 2. Review of the FDA's published listing of COVID-19 EUA's granted for SARS CoV-2 antibody testing as of 5/27/20 revealed no EUA granted for the test method outlined above. The Innovita Corona Check COVID-19 IgG/IgM rapid test method classification as of 5/27/20 was high complexity. 3. In a telephone interview with the the Director of Clinical Operations and Chief Operating Officer (COO) on 5/28/20 at approximately 1:00 PM, the inspectors asked for a description of the laboratory's QC protocols, patient test logs, and policy/procedures for the Innovita test method in use. The COO stated at approximately 1:30 PM: "We use the internal built in QC. We did not purchase extra quality control materials." 4. Review of requested patient test logs and QC records revealed 660 COVID-19 IgG/IgM patient results were reported from 5/5/20 to 5/31/20. No records of verification or negative and positive external controls on each day of patient testing was available for review. 5. Review of the laboratory's provided policy and procedures revealed no LD approved QC policy for the high complexity COVID-19 IgG/IgM test kit outlined above. The inspectors requested clarification regarding QC procedures. The COO stated in a follow up email interview on 6/17/20 at approximately 12:52 PM: "BetterMed Regency utilized the Innovita test kits from from 5/5/2020 - present. No QC materials were purchased." 6. An interview with the director of clinical operations and COO on 6/25/20, at approximately 11:00 AM, confirmed the above findings.

D6102

LABORATORY DIRECTOR RESPONSIBILITIES

CFR(s): 493.1445(e)(12)

The laboratory director must ensure that prior to testing patients' specimens, all personnel have the appropriate education and experience, receive the appropriate training for the type and complexity of the services offered, and have demonstrated that they can perform all testing operations reliably to provide and report accurate results.

This STANDARD is not met as evidenced by:
 Based on a review of the Food and Drug Administration's (FDA) Emergency Use Authorizations (EUA), Laboratory Personnel Report Form (CLIA) (CMS-209 Form), available testing personnel (TP) files, lack of documentation, and interviews, the laboratory director (LD) failed to ensure that training and competency assessments were performed for fourteen (14) of 14 TP prior to testing and reporting patient results with the Corona Check (Innovita) COVID-19 IgG/IgM Antibody test kit. Dates of record review include 5/5/20 up to 6/17/20. Findings include: 1. Review of the FDA's published listing of COVID-19 EUA granted for SARS CoV-2 antibody testing as of 5/27/20 revealed no EUA for Corona Check (Innovita) COVID-19 IgG/IgM Antibody test method. In addition, test method classification was high complexity. The COO stated in a follow up email interview on 6/17/20 at approximately 12:52 PM that the "kit was in use from 5/5/20 to present." 2. Review of the laboratory's CLIA CMS-209 Form revealed 14 TP identified as performing COVID-19 IgG/IgM patient testing with the Corona Check (Innovita) COVID-19 IgG/IgM Antibody test method. 3. Review of the available laboratory personnel records revealed lack of documentation of training and competency assessment(s) for all 14 TP (A-N) prior to use of aforementioned kit. The inspector requested to review documentation and training /competency assessments for the 14 TP. No records were available for review. See Personnel Code Sheet attached. 4. An interview with the Director of Clinical Operations and the COO on 6/25/20 at approximately 11 AM confirmed the findings.

D6168

TESTING PERSONNEL
 CFR(s): 493.1487

The laboratory has a sufficient number of individuals who meet the qualification requirements of 493.1489 of this subpart to perform the functions specified in 493.1495 of this subpart for the volume and complexity of testing performed.

This CONDITION is not met as evidenced by:
 Based on a review of the Food and Drug Administration's (FDA) Emergency Use Authorizations (EUA), Laboratory Personnel Report Form (CLIA) (CMS-209 Form), available testing personnel (TP) files, lack of documentation, and interviews, the laboratory failed to ensure that thirteen (13) of 14 TP qualified to perform high complexity testing from May 5, 2020 and up to June 17, 2020. See D6171.

D6171

TESTING PERSONNEL QUALIFICATIONS
 CFR(s): 493.1489(b)

(b) Meet one of the following requirements: (b)(1) Be a doctor of medicine, doctor of osteopathy, or doctor of podiatric medicine licensed to practice medicine, osteopathy, or podiatry in the State in which the laboratory is located or have earned a doctoral, master's or bachelor's degree in a chemical, physical, biological or clinical laboratory science, or medical technology from an accredited institution; (b)(2)(i) Have earned an associate degree in a laboratory science, or medical laboratory technology from an accredited institution or-- (b)(2)(ii) Have education and training equivalent to that specified in paragraph (b)(2)(i) of this section that includes-- (b)(2)(ii)(A) At least 60 semester hours, or equivalent, from an accredited institution that, at a minimum, include either-- (b)(2)(ii)(A)(1) 24 semester hours of medical laboratory technology courses; or (b)(2)(ii)(A)(2) 24 semester hours of science courses that include-- (b)(2)(ii)(A)(2)(i) Six semester hours of chemistry; (b)(2)(ii)(A)(2)(ii) Six semester hours of

biology; and (b)(2)(ii)(A)(2)(iii) Twelve semester hours of chemistry, biology, or medical laboratory technology in any combination; and (b)(2)(ii)(B) Have laboratory training that includes either of the following: (b)(2)(ii)(B)(1) Completion of a clinical laboratory training program approved or accredited by the ABHES, the CAHEA, or other organization approved by HHS. (This training may be included in the 60 semester hours listed in paragraph (b)(2)(ii)(A) of this section.) (b)(2)(ii)(B)(2) At least 3 months documented laboratory training in each specialty in which the individual performs high complexity testing. (b)(3) Have previously qualified or could have qualified as a technologist under 493.1491 on or before February 28, 1992; (b)(4) On or before April 24, 1995 be a high school graduate or equivalent and have either-- (b)(4)(i) Graduated from a medical laboratory or clinical laboratory training program approved or accredited by ABHES, CAHEA, or other organization approved by HHS; or (b)(4)(ii) Successfully completed an official U.S. military medical laboratory procedures training course of at least 50 weeks duration and have held the military enlisted occupational specialty of Medical Laboratory Specialist (Laboratory Technician); (b)(5)(i) Until September 1, 1997-- (b)(5)(i)(A) Have earned a high school diploma or equivalent; and (b)(5)(i)(B) Have documentation of training appropriate for the testing performed before analyzing patient specimens. Such training must ensure that the individual has-- (b)(5)(i)(B)(1) The skills required for proper specimen collection, including patient preparation, if applicable, labeling, handling, preservation or fixation, processing or preparation, transportation and storage of specimens; (b)(5)(i)(B)(2) The skills required for implementing all standard laboratory procedures; (b)(5)(i)(B)(3) The skills required for performing each test method and for proper instrument use; (b)(5)(i)(B)(4) The skills required for performing preventive maintenance, troubleshooting, and calibration procedures related to each test performed; (b)(5)(i)(B)(5) A working knowledge of reagent stability and storage; (b)(5)(i)(B)(6) The skills required to implement the quality control policies and procedures of the laboratory; (b)(5)(i)(B)(7) An awareness of the factors that influence test results; and (b)(5)(i)(B)(8) The skills required to assess and verify the validity of patient test results through the evaluation of quality control values before reporting patient test results; and (b)(5)(i)(B)(8)(ii) As of September 1, 1997, be qualified under 493.1489(b)(1), (b)(2), or (b)(4), except for those individuals qualified under paragraph (b)(5)(i) of this section who were performing high complexity testing on or before April 24, 1995; (b)(6) For blood gas analysis-- (b)(6)(i) Be qualified under 493.1489(b)(1), (b)(2), (b)(3), (b)(4), or (b)(5); (b)(6)(ii) Have earned a bachelor's degree in respiratory therapy or cardiovascular technology from an accredited institution; or (b)(6)(iii) Have earned an associate degree related to pulmonary function from an accredited institution; or (b)(7) For histopathology, meet the qualifications of 493.1449 (b) or (l) to perform tissue examinations.

This STANDARD is not met as evidenced by:

Based on a review of the Food and Drug Administration's (FDA) Emergency Use Authorizations (EUA), Laboratory Personnel Report Form (CLIA) (CMS-209 Form), available testing personnel (TP) files, lack of documentation, and interviews, the laboratory failed to ensure that thirteen (13) of 14 TP qualified to perform high complexity testing from May 5, 2020 and up to June 17, 2020. Findings include: 1. Review of the FDA's published listing of COVID-19 EUA granted for SARS CoV-2 antibody testing as of 5/27/20 revealed no EUA for the Corona Check (Innovita) test method. In addition, test method classification was high complexity. The COO stated in a follow up email interview on 6/17/20 at approximately 12:52 PM that the "kit was in use from 5/5/20 to present." 2. Review of the laboratory's CMS 209 form revealed 14 TP identified as performing COVID-19 IgG/IgM patient testing during the

timeframe of 5/5/20 to 5/29/20. 3. Review of the available laboratory personnel records revealed a lack of documentation of education requirements for high complexity testing for TP A-I, and K-N (total of 13 TP). See Personnel Code Sheet attached. 4. An interview with the Director of Clinical Operations and the COO on June 25, 2020 at approximately 11 AM confirmed the findings.