

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 50D0669238	(X3) Date Survey Completed 05/24/2018
Name of Provider or Supplier David C Wynecoop Memorial Clinic	Street Address, City, State 6203 Agency Loop Rd, Wellpinit, WA	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5437	<p>CALIBRATION AND CALIBRATION VERIFICATION CFR(s): 493.1255(a)</p> <p>Unless otherwise specified in this subpart, for each applicable test system the laboratory must perform and document calibration procedures-- (1) Following the manufacturer's test system instructions, using calibration materials provided or specified, and with at least the frequency recommended by the manufacturer; (2) Using the criteria verified or established by the laboratory as specified in 493.1253(b) (3)-- (2)(i) Using calibration materials appropriate for the test system and, if possible, traceable to a reference method or reference material of known value; and (2)(ii) Including the number, type, and concentration of calibration materials, as well as acceptable limits for and the frequency of calibration; and (3) Whenever calibration verification fails to meet the laboratory's acceptable limits for calibration verification.</p> <p>This STANDARD is not met as evidenced by: A review of chemistry calibration records and interview of laboratory personnel revealed that the laboratory failed to document the required calibrations for the Reichert Unistat Bilirubinometer. Findings: 1. The laboratory was unable to produce documentation, during the survey, of the manufacturer required every 6 months calibration of the Reichert Unistat Bilirubinometer for 2017 & 2018. 2. An interview with the laboratory manager on 05/24/2018 at 1:00 PM confirmed the lack of calibration documentation.</p>
D5439	<p>CALIBRATION AND CALIBRATION VERIFICATION CFR(s): 493.1255(b)</p> <p>Unless otherwise specified in this subpart, for each applicable test system the laboratory must do the following: Perform and document calibration verification procedure - (b)(1) Following the manufacturer's calibration verification instructions; (b)(2) Using the criteria verified or established by the laboratory under 493.1253(b)(3)</p>

-- (b)(2)(i) Including the number, type, and concentration of the materials, as well as acceptable limits for calibration verification; and (b)(2)(ii) Including at least a minimal (or zero) value, a mid-point value, and a maximum value near the upper limit of the range to verify the laboratory's reportable range of test results for the test system; and (b)(3) At least once every 6 months and whenever any of the following occur: (b)(3)(i) A complete change of reagents for a procedure is introduced, unless the laboratory can demonstrate that changing reagent lot numbers does not affect the range used to report patient test results, and control values are not adversely affected by reagent lot number changes. (b)(3)(ii) There is major preventive maintenance or replacement of critical parts that may influence test performance. (b)(3)(iii) Control materials reflect an unusual trend or shift, or are outside of the laboratory's acceptable limits, and other means of assessing and correcting unacceptable control values fail to identify and correct the problem. (b)(3)(iv) The laboratory's established schedule for verifying the reportable range for patient test results requires more frequent calibration verification.

This STANDARD is not met as evidenced by:

A review of chemistry records and interview of laboratory personnel revealed that the laboratory failed to document the required every 6 months calibration verification of the Reichert Unistat Bilirubinometer. Findings: 1. No documentation was available for review, during the survey, for the every 6 months calibration verification on the Reichert Unistat Bilirubinometer for 2017 and 2018. 2. An interview of the laboratory manager on 05/24/2018 at 1:00 PM confirmed the lack of calibration verification documentation.