

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b> 51D0236838	<b>(X3) Date Survey Completed</b> 12/14/2021
<b>Name of Provider or Supplier</b> Manchin Clinic	<b>Street Address, City, State</b> 100 Main Street, Farmington, WV	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D0000</b>	An announced, on site, recertification survey was conducted at Manchin Clinic on December 14, 2021, by the West Virginia Office of Laboratory Services. The laboratory was surveyed to assess compliance with the Federal Clinical Laboratory Improvement Amendment (CLIA) regulations under 42 CFR 493. Specific deficiencies are explained below.
<b>D2121</b>	<p>HEMATOLOGY CFR(s): 493.851(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview the laboratory failed to attain at least an 80% for the analytes 0785 HCT and 0815 Platelets in one of two Hematology 2021 proficiency testing (PT) events. Findings: 1. Review of American Proficiency Institute (API) PT records identified the following unsatisfactory analyte performances in the 2021 Hematology 1st event: 40% for analyte 0785 HCT 60% for analyte 0815 Platelets 2. During the exit interview, 12/14/2021 at approximately 1:00 PM, the technical consultant confirmed the findings.</p>