

| | | |
|--|--|---|
| Statement of Deficiencies | (X1) Provider/Supplier/CLIA Identification Number 51D0877096 | (X3) Date Survey Completed 11/12/2025 |
| Name of Provider or Supplier Camc Cancer Center | Street Address, City, State 3415 Maccorkle Avenue Se, Charleston, WV | |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. | | |

| (X4) ID Prefix Tag | Summary Statement of Deficiencies |
|---------------------------|--|
| D0000 | A routine recertification survey was conducted at Charleston Area Medical Center Cancer Center on November 12, 2025, by the West Virginia Office of Laboratory Services. The laboratory was assessed for compliance with the CLIA regulations under 42 CFR 493, Requirements for Laboratories. Specific deficiencies cited are explained below. |
| D5211 | <p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of American Proficiency Institute (API) proficiency testing (PT) records, lack of documentation, and interview with the laboratory director (LD) the laboratory failed to document the evaluation of a less than 100% performance for hematocrit in the first PT event of 2025 in hematology. Findings: 1. Review of two 2025 API PT evaluation reports for hematology revealed an 80% performance for hematocrit in hematology event 1 of 2025. PT Specimen COU-05 had a reported result outside the acceptable range and scored by API as unacceptable. 2. No documented evaluation of the less than 100% performance for hematocrit could be located. 3. During an interview on 11/12/25 at 10:40 AM, the laboratory director (LD) verified that no documented evaluation of the 80% performance for hematocrit could be located.</p> |