

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 51D1100953	(X3) Date Survey Completed 04/26/2021
Name of Provider or Supplier Renal Consultants Pllc	Street Address, City, State 24 Maccorkle Avenue Sw Suite 201, South Charleston, WV	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced, on site, recertification survey was conducted at Renal Consultants PLLC on April 26, 2021, by the West Virginia Office of Laboratory Services. The laboratory was surveyed to assess compliance with the Federal Clinical Laboratory Improvement Amendment (CLIA) regulations under 42 CFR 493. Specific deficiencies are explained below.
D2087	<p>ROUTINE CHEMISTRY CFR(s): 493.841(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on record review and interview, the laboratory failed to achieve a satisfactory score for analyte 0585 TSH in 1 of 3 American Proficiency Institute (API) 2020 testing events. Findings: 1. API record review identified a score of 60% for analyte 0585 TSH in the 2nd testing event of 2020. 2. CASPER 155D Report confirmed the API score of 60%. 3. The laboratory manager confirmed the finding 4/26/2021 at approximately 11:10 AM.</p>