Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number**	**(X3) Date Survey Completed**
	51D2010705	01/12/2023
Name of Provider or Supplier	**Street Address, City, State**	
Derm One Pllc	4130 Robert C Byrd Drive, Beckley, WV	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	An announced, on site, routine recertification survey was conducted at Derm One PLLC on January 12, 2023, by the West Virginia Office of Laboratory Services. The laboratory was surveyed to assess compliance with the Federal Clinical Laboratory Improvement Amendment (CLIA) regulations under 42 CFR 493. Specific deficiencies cited are explained below.
D5217	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1) At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part. This STANDARD is not met as evidenced by: Based on record review and interview the laboratory failed to document the biannual evaluation and comparison for alternative accuracy verification of the MOHs procedure for one of two events in 2021 and two of two events in 2022. Findings: 1. Review of 2021 MOHs accuracy verification logs and records identified no documentation of the comparative evaluation between the outsourced diagnosis and the in house diagnosis for the second event in 2021. 2. Review of 2022 MOHs accuracy verification logs and records identified no documentation of the comparative evaluation between the outsourced diagnosis and the in house diagnosis for two of two events in 2022. 3. An interview with the laboratory manager, 1/12/23 at approximately 10:30 AM, confirmed the lack of documentation for the evaluation of the alternative accuracy verification process.