

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 51D2076798	(X3) Date Survey Completed 02/21/2018
Name of Provider or Supplier Coplin Clinic Laboratory	Street Address, City, State 483 Court St, Elizabeth, WV	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2087	<p>ROUTINE CHEMISTRY CFR(s): 493.841(a)</p> <p>Failure to attain a score of at least 80 percent of acceptable responses for each analyte in each testing event is unsatisfactory analyte performance for the testing event.</p> <p>This STANDARD is not met as evidenced by: Based on review of the laboratory's American Proficiency Institute (API) proficiency testing (PT) records and interview with the Technical Consultant #1 (TC1), the laboratory failed to achieve satisfactory performance for the analyte of Sodium for 2017 Chemistry-Core-2nd Event. The findings include: 1. Review of the API proficiency testing records identified the result for 2017 Chemistry-Core-2nd Event Sodium was 60%. 2. On 2/21/18 at approximately 1:30 PM, TC1 confirmed the findings.</p>