

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 51D2118832	(X3) Date Survey Completed 12/20/2023
Name of Provider or Supplier Huntington Hospitalist Group	Street Address, City, State 4540 Us Rt 60, Huntington, WV	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A routine recertification survey was conducted at Huntington Hospitalist Group on December 20, 2023, by the West Virginia Office of Laboratory Services. The laboratory was assessed for compliance with the Federal Clinical Laboratory Improvement Amendments (CLIA) regulations under 42 CFR 493. Specific deficiencies cited are explained below.
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on record review, lack of documentation, and interview the laboratory failed to verify the accuracy of 11 of the 29 analytes tested on the AB SCIEX toxicology analyzer in 2023. Findings: 1. Review of proficiency testing (PT) records for 2023 identified 11 analytes that are not included in Subpart I and not enrolled in commercial PT. 2. No documentation for the verification of accuracy for the 11 analytes on the Sciex analyzer could be located for 2023. 3. An exit interview with the laboratory director on 12/20/23 at 11:30 AM, confirmed the findings.</p>