

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 51D2151883	(X3) Date Survey Completed 06/25/2025
Name of Provider or Supplier Chesapeake Healthcare Center, Llc	Street Address, City, State 11950 Maccorkle Avenue, Charleston, WV	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D0000	A routine recertification survey was conducted at Chesapeake Healthcare Center LLC. on June 25, 2025, by the West Virginia Office of Laboratory Services. The laboratory was assessed for compliance with the CLIA regulations at 42 CFR 493, Requirements for Laboratories. Specific deficiencies cited are explained below.
D5209	<p>PERSONNEL COMPETENCY ASSESSMENT POLICIES CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on review of laboratory policies and procedures, testing personnel files, lack of documentation, and interview with testing personnel (TP3), the laboratory failed to document the initial competency assessment (CA) for two of two laboratory testing personnel in 2024. Findings: 1. Review of laboratory policies and procedures identified "Quality Assessment Plan" stating "Employees will be authorized to run tests only after completion of training and documentation of competency at initial, 6 months, and annual intervals." 2. Review of 2024 and 2025 testing personnel CA files revealed no documentation for the initial competency of TP1 and TP2. TP1 began testing February 2024, 6 month CA 7/10/24 TP2 began testing November 2024, 6 month CA 5/20/25 3. During an interview with TP3, 6/25/25 at 9:50 AM, TP3 agreed that no documentation of TP1 and TP2 initial competency assessments could be located.</p>