

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  52D0391437	<b>(X3) Date Survey Completed</b>  08/30/2021
<b>Name of Provider or Supplier</b>  Layton Avenue Dermatology	<b>Street Address, City, State</b>  2923 W Layton Ave, Greenfield, WI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D6127</b>	<p><b>TECHNICAL SUPERVISOR RESPONSIBILITIES</b> CFR(s): 493.1451(b)(9)</p> <p>The technical supervisor is responsible for evaluating and documenting the performance of individuals responsible for high complexity testing at least semiannually during the first year the individual tests patient specimens.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of testing personnel records and interview with the practice manager, staff A, the technical supervisor did not evaluate the semiannual competency for one of one new testing personnel performing high complexity testing. Findings include: 1. Review of testing personnel records showed initial training performed in June 2020 for staff B. Further review showed no evaluation of semiannual competency for staff B. 2. Interview with staff A on August 30, 2021 at 10:05 AM confirmed the technical supervisor did not evaluate the semiannual competency for one of one new testing personnel performing high complexity testing.</p>