

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  52D0394413	<b>(X3) Date Survey Completed</b>  09/14/2023
<b>Name of Provider or Supplier</b>  Oneida Community Health Center	<b>Street Address, City, State</b>  525 Airport Dr, Oneida, WI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5775</b>	<p>COMPARISON OF TEST RESULTS CFR(s): 493.1281(a)(c)</p> <p>(a) If a laboratory performs the same test using different methodologies or instruments, or performs the same test at multiple testing sites, the laboratory must have a system that twice a year evaluates and defines the relationship between test results using the different methodologies, instruments, or testing sites. (c) The laboratory must document all test result comparison activities.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor observation of equipment in the laboratory, review of records and interview with the laboratory director, the laboratory has not evaluated the relationship between the two MedTox Urine Drug Screen analyzers used to test patient samples with Profile V cartridges. The laboratory did not complete four of four evaluations in the last two years. Findings include: 1. Observation of equipment in the laboratory on September 14, 2023, at 10:00 AM revealed two MedTox Urine Drug Screen Analyzers were available for testing. 2. Review of laboratory records showed no evidence of comparison or evaluation twice annually of the relationship of results from the two analyzers. 3. Interview with the laboratory director on September 14, 2023, at 3:45 PM confirmed the laboratory had not evaluated the relationship between test results produced by the two analyzers twice annually in the past two years.</p>