

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  52D0395521	<b>(X3) Date Survey Completed</b>  03/04/2026
<b>Name of Provider or Supplier</b>  Aspirus Kronenwetter Clinic	<b>Street Address, City, State</b>  1881 Cty Hwy Xx, Mosinee, WI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5209</b>	<p><b>PERSONNEL COMPETENCY ASSESSMENT POLICIES</b> CFR(s): 493.1235</p> <p>As specified in the personnel requirements in subpart M, the laboratory must establish and follow written policies and procedures to assess employee and, if applicable, consultant competency.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of laboratory records and interview with the laboratory director, the laboratory did not document site-specific competency for three of three testing personnel that occasionally perform testing at this laboratory. Findings include: 1. Review of laboratory personnel records showed no competence evaluations of the three testing personnel (Staff A, B, and C) who occasionally perform testing at this laboratory. 2. Interview with the laboratory director on March 4, 2026, at 9:00 AM confirmed the laboratory had not documented full competence assessments for the three occasional testing personnel at this location.</p>