

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 52D0396431	(X3) Date Survey Completed 01/22/2026
Name of Provider or Supplier Kickapoo Valley Medical Clinic-Vmh	Street Address, City, State 102 Sunset Ave, Soldiers Grove, WI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of proficiency testing and laboratory records and interview with the laboratory director, the laboratory did not verify the accuracy of two tests twice annually in two of the last two years. Findings include: 1. Review of proficiency testing records showed no evidence of participation in proficiency testing evaluation for microscopic urinalysis testing and microscopic potassium hydroxide (KOH) testing on skin specimens in 2024 and 2025. 2. Review of laboratory records showed no evidence of twice annual accuracy checks of microscopic urinalysis testing and microscopic KOH testing on skin specimens 2024 and 2025. 3. Interview with the laboratory director on January 22, 2026, at 11:50 AM confirmed the laboratory offered microscopic urinalysis testing and microscopic KOH testing on skin specimens in 2024 and 2025, and confirmed the laboratory had not verified the accuracy of the two tests twice annually in 2024 and 2025. This is a repeat deficiency, previously cited on November 28, 2023.</p>