

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 52D0396654	(X3) Date Survey Completed 03/07/2018
Name of Provider or Supplier Western Wisconsin Urology	Street Address, City, State 3217 Stein Blvd, Eau Claire, WI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5807	<p>TEST REPORT CFR(s): 493.1291(d)</p> <p>Pertinent "reference intervals" or "normal" values, as determined by the laboratory performing the tests, must be available to the authorized person who ordered the tests and, if applicable, the individual responsible for using the test results.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of patient test reports, laboratory procedures, and interview with the laboratory lead tech, the reference intervals (normal values) for urine microscopic's were not available. Findings include: 1. Review of patient test reports in the electronic medical record and on the paper "Urinalysis Report" for urine microscopic results show that the laboratory does not list the reference intervals (normal values) on the patient test reports. 2. Review of the "Urine Reporting Formats" laboratory procedure does not state what the normal values are for urine microscopic result reporting. 3. Interview with the laboratory lead tech, staff A, on March 7, 2018 at 2:00 PM confirmed that no reference intervals were available for the reported urine microscopy test results.</p>