Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 52D0396920	(X3) Date Survey Completed 02/04/2020
Name of Provider or Supplier Essentia Health Spooner Clinic	**Street Address, City, State** 1180 Chandler Dr, Spooner, WI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5217	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1) At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part. This STANDARD is not met as evidenced by: Based on surveyor review of proficiency testing records and interview with a technical consultant, the laboratory has not verified the accuracy of semi-quantitative sperm counts twice annually. Findings include: 1. Review of proficiency testing records from 2018 and 2019 showed the laboratory did not verify the accuracy of semi-quantitative sperm counts using proficiency samples. 2. Interview with staff A, a technical consultant, on February 4, 2020 at 2:00 PM confirmed the laboratory reports semi-quantitative sperm counts as part of their post-vasectomy semen analysis. Further interview confirmed the laboratory did not verify the accuracy of the test twice annually.