

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 52D0665546	(X3) Date Survey Completed 09/20/2023
Name of Provider or Supplier Prohealth Medical Group, Inc -	Street Address, City, State 210 Nw Barstow St Suite 201, Waukesha, WI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5215	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(2)</p> <p>The laboratory must verify the accuracy of any analyte, specialty or subspecialty assigned a proficiency testing score that does not reflect laboratory test performance (that is, when the proficiency testing program does not obtain the agreement required for scoring as specified in subpart I of this part, or the laboratory receives a zero score for nonparticipation, or late return or results).</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of proficiency testing (PT) records and interview with the technical consultant, the laboratory did not evaluate one of one not scored / non-consensus result in the second testing event in 2022. Finding include: 1. Review of the Wisconsin State Laboratory of Hygiene (WSLH) PT results from the miscellaneous quality assurance point of care second event (MISC QA POC2) showed sample PM-4 for urine sediment was resulted by the laboratory as bacteria. The expected result was calcium carbonate crystals. The report showed no evidence of review by the laboratory. 2. Interview with the technical consultant on September 20, 2023, at 10:00 AM confirmed the laboratory did not review the ungraded result to verify the accuracy of the reported result.</p>