

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 52D0670794	(X3) Date Survey Completed 04/26/2021
Name of Provider or Supplier Uw Eau Claire Student Health Service	Street Address, City, State Crest Wellness Center, Eau Claire, WI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5215	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(2)</p> <p>The laboratory must verify the accuracy of any analyte, specialty or subspecialty assigned a proficiency testing score that does not reflect laboratory test performance (that is, when the proficiency testing program does not obtain the agreement required for scoring as specified in subpart I of this part, or the laboratory receives a zero score for nonparticipation, or late return or results).</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of Wisconsin State Laboratory of Hygiene (WSLH) proficiency testing (PT) records and interview with the technical consultant, the laboratory did not evaluate two of two analytes that received a score of 0% for provider-performed microscopy (PPM) in event one of 2020. Findings include: 1. Review of PPM proficiency testing records for 2020-MiscQA_POC1 showed the laboratory received 0% scores for "Pinworm Preparation" and "Vaginal Wet Prep" due to a submission error when resulting the analytes to the PT provider. Further review showed no evidence the laboratory evaluated the results from the PT provider's participant summary to verify the accuracy of their responses. 2. Interview with the technical consultant on April 26, 2021 at 1:45 PM confirmed the laboratory did not evaluate two of two analytes that received a score of 0% for PPM in event one of 2020.</p>