

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  52D0716590	<b>(X3) Date Survey Completed</b>  04/12/2021
<b>Name of Provider or Supplier</b>  Uw La Crosse Student Health Center	<b>Street Address, City, State</b>  1300 Badger St, La Crosse, WI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5417</b>	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor observation of testing reagents, review of patient reports and interview with testing personnel, staff A, the laboratory used potassium hydroxide (KOH) past the expiration date on the bottle. Findings include: 1. Observation of the KOH reagent available on April 12, 2021 at 10:15 AM showed the expiration date printed on the bottle was January 31, 2021 (2021-01-31). 2. Review of patient test reports from February and March 2021 showed the laboratory performed four patient KOH tests. 3. Interview with staff A on April 12, 2021 at 10:15 AM confirmed testing personnel performed patient testing using the observed bottle of KOH past the expiration date of the reagent.</p>