

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 52D0867947	(X3) Date Survey Completed 03/27/2025
Name of Provider or Supplier Uw Health Central Laboratory	Street Address, City, State 20 S Park St Suite 465, Madison, WI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5417	<p>TEST SYSTEMS, EQUIPMENT, INSTRUMENTS, REAGENT CFR(s): 493.1252(d)</p> <p>(d) Reagents, solutions, culture media, control materials, calibration materials, and other supplies must not be used when they have exceeded their expiration date, have deteriorated, or are of substandard quality.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of laboratory records and interview with a technical consultant, staff A, the laboratory used expired reagents to perform potassium hydroxide (KOH) testing for three of three patient tests between February 2-12, 2025. Findings include: 1. Review of laboratory records showed KOH Lot 240806-020 expired on February 2, 2025, and KOH Lot 241227-038 was opened on February 12, 2025. 2. Review of patient testing in the laboratory information system (LIS) showed three KOH patient tests had been performed between February 2, 2025, and February 12, 2025. 3. Interview with staff A on March 27, 2025, at 1:00 PM confirmed the laboratory used expired reagents to perform KOH testing. This is a repeat deficiency from August 1, 2023.</p>