

<b>Statement of Deficiencies</b>	<b>(X1) Provider/Supplier/CLIA Identification Number</b>  52D0869645	<b>(X3) Date Survey Completed</b>  02/13/2024
<b>Name of Provider or Supplier</b>  Allergy Immunology Diagnostic Lab Center	<b>Street Address, City, State</b>  8701 Watertown Plank Rd, Milwaukee, WI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

<b>(X4) ID Prefix Tag</b>	<b>Summary Statement of Deficiencies</b>
<b>D5217</b>	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(c)(1)</p> <p>At least twice annually, the laboratory must verify the accuracy of any test or procedure it performs that is not included in subpart I of this part.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of laboratory records and interview with testing personnel, staff A, the laboratory did not verify the accuracy of Enzyme Linked Immunosorbent Assay (ELISA) testing for two of two events in 2023 and did not verify the accuracy of immunodiffusion testing for one of two events in 2023. Findings include: 1. Review of the laboratory test menu showed the laboratory performs ELISA and immunodiffusion testing. 2. Review of laboratory records showed no evidence of twice annual accuracy checks of ELISA testing in 2023. Further review showed only one event for immunodiffusion testing in 2023. 3. Interview with the staff A on February 13, 2024, at 10:15 AM confirmed the laboratory had not verified the accuracy of the laboratory testing twice annually in 2023.</p>