

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 52D0884857	(X3) Date Survey Completed 03/03/2023
Name of Provider or Supplier Aurora Health Care Metro Inc - Metro East -	Street Address, City, State 3003 W Good Hope Rd, 2nd Floor Oncology, Milwaukee, WI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5213	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(1)</p> <p>The laboratory must verify the accuracy of any analyte or subspecialty without analytes listed in subpart I of this part that is not evaluated or scored by a CMS-approved proficiency testing program.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of proficiency testing (PT) records and interview with the technical supervisor, the laboratory did not evaluate the accuracy for ten of fifteen "Educational-see Event Note" PT results for nucleated red blood cells (NRBC) in 2022. Findings include: 1. Review of Wisconsin State Laboratory of Hygiene (WSLH) PT records showed WSLH entered notes for the NRBC analyte stating "Educational-see Event note" for fifteen of fifteen results in 2022. Further review showed the laboratory did not evaluate the accuracy of the NRBC PT results for five of five results for event one and five of five results for event three in 2022. 2. Interview with the technical supervisor on March 3, 2023 at 9:50 AM confirmed the laboratory did not evaluate the accuracy for ten of fifteen "Educational-see Event Note" PT results for nucleated red blood cells (NRBC) in 2022.</p>