

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 52D0968924	(X3) Date Survey Completed 12/14/2021
Name of Provider or Supplier Consultants Laboratory Of Wi, Llc	Street Address, City, State 145 N Main St, Fond Du Lac, WI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D6046	<p>TECHNICAL CONSULTANT RESPONSIBILITIES CFR(s): 493.1413(b)(8)</p> <p>(b) The technical consultant is responsible for-- (b)(8) Evaluating the competency of all testing personnel and assuring that the staff maintain their competency to perform test procedures and report test results promptly, accurately and proficiently.</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of testing personnel competency assessments and interview with a technical consultant, staff A, the technical consultant did not complete a semi-annual competency evaluation for one of one new testing personnel, staff B. Findings include: 1. Review of testing personnel competency assessment records revealed competency evaluation for Staff B was not documented twice during the first year. 2. Interview with staff A on December 14, 2021 at 9:10 AM confirmed competency evaluation for Staff B was not documented twice during the first year of testing. This is a repeat deficiency from November 18, 2019.</p>