Statement of Deficiencies	**(X1) Provider/Supplier/CLIA Identification Number**	**(X3) Date Survey Completed**
	52D2126222	02/13/2019
Name of Provider or Supplier	**Street Address, City, State**	
Rprd Diagnostics, Llc	1225 Discovery Parkway, Suite 260, Wauwatosa, WI	

For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5213	EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(1) The laboratory must verify the accuracy of any analyte or subspecialty without analytes listed in subpart I of this part that is not evaluated or scored by a CMS-approved proficiency testing program. This STANDARD is not met as evidenced by: Based on surveyor review of proficiency testing reports from the College of American Pathologists (CAP) and interview with the general supervisor, the laboratory did not evaluate the accuracy of the results that the proficiency testing provider did not grade in the first PGX event in 2018. Findings include: 1. Review of the proficiency testing reports from CAP for the first PGX event in 2018 showed interpretation and clinical management results were not graded. There is no indication that the laboratory reviewed these results. 2. Interview with the general supervisor on February 13, 2019 at 11:30 AM confirmed the laboratory had not reviewed or verified the accuracy of results that were not graded.