

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 52D2133898	(X3) Date Survey Completed 03/31/2022
Name of Provider or Supplier Fmccp Tosa Health Center	Street Address, City, State 1155 N Mayfair Road, Wauwatosa, WI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on surveyor review of the federal Certification and Survey Provider Enhancement Reports (CASPER) Proficiency Testing (PT) and the American Proficiency Institute (API) Proficiency Testing (PT) records, the laboratory failed to successfully participate in PT for Chloride in the Specialty of Chemistry for events 2021-2 and 2022-1. See D2096.</p>
D2096	<p>ROUTINE CHEMISTRY CFR(s): 493.841(f)</p>

Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on surveyor review of the federal Certification and Survey Provider Enhancement Reports (CASPER) Proficiency Testing (PT) and the American Proficiency Institute (API) Proficiency Testing (PT) records, the laboratory failed to successfully participate in PT for Chloride in the Specialty of Chemistry for events 2021-2 and 2022-1. Findings include: 1. Review of the PT records in the CASPER reporting system showed the laboratory failed two of three PT events for Chloride in the Specialty of Chemistry: Analyte / 2021-2 / 2022-1 CL / 60% /60% 2. Review of the API PT evaluation reports on March 31, 2022, confirmed the failed PT scores for Chloride in the Specialty of Chemistry. This results in failure to achieve an overall testing event score of satisfactory performance for two out of three testing events which is an unsuccessful performance.