

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 52D2181806	(X3) Date Survey Completed 04/01/2021
Name of Provider or Supplier Aurora Health Center - Pleasant Prairie	Street Address, City, State 12500 Aurora Drive, Pleasant Prairie, WI	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5215	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(b)(2)</p> <p>The laboratory must verify the accuracy of any analyte, specialty or subspecialty assigned a proficiency testing score that does not reflect laboratory test performance (that is, when the proficiency testing program does not obtain the agreement required for scoring as specified in subpart I of this part, or the laboratory receives a zero score for nonparticipation, or late return or results).</p> <p>This STANDARD is not met as evidenced by: Based on surveyor review of Wisconsin State Laboratory of Hygiene (WSLH) proficiency testing (PT) records and interview with the technical consultant, the laboratory did not evaluate the accuracy for two of two "Not scored-non-consensus" proficiency testing results in 2020 and 2021. Findings include: 1. Review of WSLH PT records showed the laboratory did not evaluate the accuracy of the following PT results in 2020 and 2021: a. 2020-Chem/Endo/Tx3: One of one "Not scored-non-consensus" hCG, serum result. b. 2021-Chem/Endo/Tx1: One of one "Not scored-non-consensus" Lipase result. 2. Interview with the technical consultant on April 1, 2021 at 9:28 AM, confirmed the laboratory did not evaluate the accuracy for two of two "Not-scored-non-consensus" PT results in 2020 and 2021.</p>