

|  |  |   |
|--|--|---|
| <b>Statement of Deficiencies</b>   | <b>(X1) Provider/Supplier/CLIA Identification Number</b><br>52D2249644 | <b>(X3) Date Survey Completed</b><br>11/02/2022 |
| <b>Name of Provider or Supplier</b><br>Student Health Centre - Uw Stout Student Clinic                                     | <b>Street Address, City, State</b><br>103 1st Ave W, Menomonie, WI     |   |
| For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency. |  |   |

| <b>(X4) ID Prefix Tag</b> | <b>Summary Statement of Deficiencies</b>  |
|---------------------------|---|
| <b>D5807</b>              | <p>TEST REPORT<br/>CFR(s): 493.1291(d)</p> <p>Pertinent "reference intervals" or "normal" values, as determined by the laboratory performing the tests, must be available to the authorized person who ordered the tests and, if applicable, the individual responsible for using the test results.</p> <p>This STANDARD is not met as evidenced by:<br/>Based on survey review of a patient's Complete Blood Count (CBC) test report and laboratory procedures and interview with the technical supervisor, the reference range shown on the patient report was not the same as the approved reference range in the procedure. Findings include: 1. Review of the reference range of the CBC patient test reports showed the ranges for hematocrit (HCT) and Mean Corpuscular Volume (MPV) are: Patient/age/HCT %/MCV fL One/20 year-old male/40.0-52.0/83.0-94.0 Two/22-year old female/30.0-52.0/83.0-99.0 2. Review of the "Horiba ABX Micros 60" procedure showed the approved reference ranges for are: Patient/age/HCT % /MCV fL One/20-year old male/39.0-51.0/80.0-98.0 Two/22-year old female/32.0-42.0/80.0-98.0 3. Interview with the technical supervisor on November 2, 2022, at 10: 50 AM confirmed the reference range on the patient report was not consistent with the approved reference range in the procedure.</p> |