

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 53D0519862	(X3) Date Survey Completed 02/26/2021
Name of Provider or Supplier So Big Horn Co Hospital	Street Address, City, State 388 Us Hwy 20 South, Basin, WY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: Based on American Proficiency Institute (API) report review, Centers for Medicare and Medicaid Services, (CMS) Casper 155 report review, and telephone interview conducted on 02/26/2021, the laboratory failed to successfully participate in proficiency testing for Chloride (CL) in two out of three events (the 2nd event of 2020 and the 1st event of 2021). (See D2096.)</p>
D2096	<p>ROUTINE CHEMISTRY CFR(s): 493.841(f)</p>

Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

Based on American Proficiency Institute (API) report review, CMS CASPER 155 report review, and telephone confirmation of test results, the laboratory failed to successfully participate in 2 out of 3 API Chemistry core proficiency testing events for Chloride, (CL). Findings include: 1. API summary evaluation reports review for the second Chemistry event of 2020 and the first Chemistry event of 2021 included summary scores of 40% Chloride scores for the 2nd event of 2020 and 60% for the 1st API Chemistry testing event of 2021. 2. Casper 155 report for the 2nd event of 2020 review included Chloride scores of 40% for the 2nd event of 2020. 3. In a telephone interview on 02/26/2021, the laboratory technical supervisor verified failure of the 2nd event of 2020 and the 1st event of 2021 making 2 out of 3 Chloride proficiency failures.