

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 53D0519922	(X3) Date Survey Completed 10/14/2019
Name of Provider or Supplier Lander Medical Clinic	Street Address, City, State 745 Buena Vista Dr, Lander, WY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	<p>SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c)</p> <p>(a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history.</p> <p>This CONDITION is not met as evidenced by: . Based on American Proficiency Institute (API) report review, Centers for Medicare and Medicaid Services (CMS) Casper 155 report review, and E-Mail interview conducted on 10/07/2019, and telephone interview on 10/15/2019, the laboratory failed to successfully participate in proficiency testing for Chloride (CL) in two consecutive testing events (the 2nd and the 3rd events of 2019). (See D2096.) .</p>
D2096	<p>ROUTINE CHEMISTRY CFR(s): 493.841(f)</p>

Failure to achieve satisfactory performance for the same analyte or test in two consecutive testing events or two out of three consecutive testing events is unsuccessful performance.

This STANDARD is not met as evidenced by:

. Based on American Proficiency Institute (API) report review, CMS CASPER 153 and 155 reports review, and E-mail confirmation of test results, the laboratory failed to successfully participate in 2 consecutive proficiency testing events for Chloride, (CL) (API events 2 and 3 of 2019). Findings include: 1. API summary evaluation reports review for the second and third Chemistry events of 2018 included summary scores of 0% Chloride scores for the for the 2nd and 3rd American Proficiency testing agency events of 2019. 2. Casper reports for 2 out of 3 events reviewed included Chloride scores of 0% for the 2nd and 3rd events of 2019. 3. In an E-mail interview on 10/07/2019 the laboratory technical consultant stated the laboratory failed the 2nd and 3rd 2019 events making 2 consecutive Chloride failures.