Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 53D0666137	(X3) Date Survey Completed 04/29/2018
Name of Provider or Supplier Star Valley Medical Center	**Street Address, City, State** 901 Adams Street, Afton, WY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D2016	SUCCESSFUL PARTICIPATION CFR(s): 493.803(a)(b)(c) (a) Each laboratory performing nonwaived testing must successfully participate in a proficiency testing program approved by CMS, if applicable, as described in subpart I of this part for each specialty, subspecialty, and analyte or test in which the laboratory is certified under CLIA. (b) Except as specified in paragraph (c) of this section, if a laboratory fails to participate successfully in proficiency testing for a given specialty, subspecialty, analyte or test, as defined in this section, or fails to take remedial action when an individual fails gynecologic cytology, CMS imposes sanctions, as specified in subpart R of this part. (c) If a laboratory fails to perform successfully in a CMS-approved proficiency testing program, for the initial unsuccessful performance, CMS may direct the laboratory to undertake training of its personnel or to obtain technical assistance, or both, rather than imposing alternative or principle sanctions except when one or more of the following conditions exists: (1) There is immediate jeopardy to patient health and safety. (2) The laboratory fails to provide CMS or a CMS agent with satisfactory evidence that it has taken steps to correct the problem identified by the unsuccessful proficiency testing performance. (3) The laboratory has a poor compliance history. This CONDITION is not met as evidenced by: Based on American Proficiency Institute (API) report review, Centers for Medicare and Medicaid Services, (CMS) Casper 155 report review, and E-Mail interview conducted on 04/29/2018, the failed to successfully participate in proficiency testing for the endocrinology analyte Free Thyroxine (FT4) in two consecutive testing events (the 3rd event of 2017 and the 1st event of 2018). See D2017.
D2017	REINSTATEMENT OF NONWAIVED LABORATORIES CFR(s): 493.807(a)(b)

	(a) If a laboratory's certificate is suspended or limited or its Medicare or Medicaid approval is cancelled or its Medicare or Medicaid payments are suspended because it fails to participate successfully in proficiency testing for one or more specialties, subspecialties, analyte or test, or voluntarily withdraws its certification under CLIA for the failed specialty, subspecialty, or analyte, the laboratory must then demonstrate sustained satisfactory performance on two consecutive proficiency testing events, one of which may be on site, before CMS will consider it for reinstatement for certification and Medicare or Medicaid approval in that specialty, subspecialty, analyte or test. (b) The cancellation period for Medicare and Medicaid approval or period for suspension of Medicare or Medicaid payments or suspension or limitation of certification under CLIA for the failed specialty, subspecialty, or analyte or test is for a period of not less than six months from the date of cancellation, limitation or suspension of the CLIA certificate. This CONDITION is not met as evidenced by: Based onAmerican Proficinecy Institute (API) report review, CMS CASPER 153 and 155 reports review, and E-mail confirmation of test results, the laboratory failed to successfully participate in 2 consecutive proficiency testing events for Free Thyroxine (Free T4) (API event 1 of 2018 and event 2 of 2017). Findings include: 1. API summary evaluation reports review for the first Chemistry and Endocrinology even of 2018 and second Endocrinology proficiency testing event of 2017 included summary scores of 60% for score for the Free T4 analyte for the 3rd American Proficiency testing agency event of 2017. 2. Casper reports for 2 out of 3 events reviewed included Endocrinology scores of 0% for the 1st and 2nd events of 2017 . 3. In an E-mail interview conducted on 04/ 18/2018 the laboratory technical supervisor stated they had failed 2 events in a row.