

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 53D0675719	(X3) Date Survey Completed 09/21/2021
Name of Provider or Supplier Veterans' Home Of Wyoming	Street Address, City, State 700 Veterans' Lane, Buffalo, WY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D1002	<p>REPORTING OF SARS-CoV-2 TEST RESULTS</p> <p>During the Public Health Emergency, as defined in 400.200 of this chapter, each laboratory that performs a test that is intended to detect SARS-CoV-2 or to diagnose a possible case of COVID-19 (hereinafter referred to as a "SARS-CoV-2 test") must report SARS-CoV-2 test results to the Secretary in such form and manner, and at such timing and frequency, as the Secretary may prescribe.</p> <p>This CONDITION is not met as evidenced by: Based on record review, review of the laboratory's policy and procedure, and staff interview, the laboratory failed to report 840 negative SARS-CoV-2 resident and staff test results, as required, for 32 weeks of testing reviewed (2/4/21 to 9/15/21). The findings were: 1. Review of the laboratory's documentation showed 841 resident and staff tests were performed using the Abbott Binax Now COVID-19 test system from 2/4/21 to 9/15/21. There was no evidence the 840 negative SARS-CoV-2 resident and staff results had been reported to the State Public Health Laboratory. 2. Review of the laboratory's procedure titled "Testing/Reporting for COVID-19" last revised 11/6/20 showed no procedure for reporting negative resident and staff test results. 3. Interview with the laboratory director on 9/21/21 at 9:30 AM confirmed the negative SARS-CoV-2 resident and staff test results had not been reported to the State Public Health Laboratory.</p>