

Statement of Deficiencies	(X1) Provider/Supplier/CLIA Identification Number 53D1049790	(X3) Date Survey Completed 09/18/2025
Name of Provider or Supplier Rocky Mountain Oncology	Street Address, City, State 6501 East 2nd Street, Casper, WY	
For information on the provider's plan to correct this deficiency, please contact the provider or the state survey agency.		

(X4) ID Prefix Tag	Summary Statement of Deficiencies
D5211	<p>EVALUATION OF PROFICIENCY TESTING PERFORMANCE CFR(s): 493.1236(a)</p> <p>The laboratory must review and evaluate the results obtained on proficiency testing performed as specified in subpart H of this part.</p> <p>This STANDARD is not met as evidenced by: Based on review of the American Proficiency Institute (API) proficiency testing records, lack of documentation, and staff interview, the laboratory failed to review and evaluate proficiency testing results for 2 of 11 testing events reviewed from August 2023 through August 2025. The findings were: 1. Review of the API proficiency testing (PT) report failed to include documentation the laboratory had evaluated test scores of less than 100%. The following concerns were identified: a. Review of the 2025 API chemistry core event #1 results showed the laboratory scored an 80% on glucose and magnesium. There was no documentation the laboratory had evaluated the proficiency testing results. b. Review of the 2025 API chemistry core event #2 results showed the laboratory scored an 80% on alkaline phosphatase, iron, and magnesium. There was no documentation the laboratory had evaluated the proficiency testing results. 2. Interview with the technical consultant on 9/17/25 at 3:07 PM confirmed the laboratory had failed to evaluate the reasons for the 80% scores.</p>